


axiomX3®

Technical and scientific review



KEY CHAPTERS

- #1 - Bone is gold..... 4
- #2 - Outstanding resistance..... 8
- #3 - Efficient osseointegration..... 10
- #4 - Towards shorter treatments..... 13

“**AXIOM X3[®] DESIGN ALLOWS FOR A REDUCTION IN THE NUMBER OF SURGICAL STEPS COMPARED TO AXIOM REG AND SO THE SAVING OF BONE, WHILE TARGETING FIRM PRIMARY STABILITY IN VARIOUS CLINICAL SITUATIONS.³¹”**

Entering the development of a new dental implant nowadays is a real challenge for a manufacturer. The current clinical results are extremely satisfying, with a broad scope of indications covered by our Axiom[®] range. The system is firmly relying on evidence-based assets with a strong material, a clean and performing surface treatment and a single conical connection. But still, there was a clear incentive to start Axiom X3[®] development: the exchanges with our customers. Discussing with practitioners, observing their practice on a daily basis not only reveals their commitment to restore patient smiles with the highest degree of predictability but also their wish to provide efficient treatment workflows. Relying on scientific foundations on one hand, tailoring the Axiom X3 solution to your needs on the other hand, thanks to technical innovation.

We deeply studied the bone-implant interaction at implant insertion, which resulted in the distinctive design of Axiom X3[®] threads, now protected by two patents. This design allows for a reduction in the number of drilling and tapping steps compared to Axiom REG, and so the saving of bone, while targeting a firm primary stability in various clinical situations³¹. The significance of these improvements was verified by a large survey performed at the European level [1]. Feedback from 63 practitioners from 9 countries on 706 implant placements were compiled, including immediate placement and/or loading in 32% of the cases. A very high satisfaction level of 89% was obtained regarding the implant stability, and a possible reduction in the number of drilling steps by 25% was confirmed compared to Axiom[®] REG.



Nicolas COURTOIS

Head of Research
& Clinical Affairs

INTERNATIONAL PRODUCT LAUNCH SURVEY


A LARGE GROUP OF EUROPEAN PRACTITIONERS WERE SURVEYED TO MEASURE THE IMPACT OF AXIOM X3® DESIGN IN TERMS OF NUMBER OF SURGICAL STEPS AND PRIMARY STABILITY.¹




UNIVERSAL BONE ANCHORAGE

A VERY HIGH SATISFACTION LEVEL WAS OBTAINED REGARDING THE IMPLANT STABILITY IN VARIOUS CLINICAL SITUATIONS.¹

DISTRIBUTION OF IMPLANTS PLACED



PERCEIVED PRIMARY STABILITY



A CHALLENGING EQUATION TO SOLVE

Dental implants are used in very different clinical situations, resulting in the need to adapt the preparation protocol and/or to resort to the use of different implant designs. The latter choice forces the user to store a substantial stock for each range, and sometimes even to deal with different prosthetic systems and/or different instruments.


Most often, in hard bone a tap is required to prepare the implant bed. This step is time consuming and a potential source for error when setting the motor speed or direction. The aim of the tap is to create a female thread matching the shape of the implant by removing some bone volume. This implies the removal of a large bone volume exceeding the implant size.

As an alternative to the use of a tap, some manufacturers indicate the use of a final drill whose size is very close to the implant diameter (0.1 or 0.2 mm larger), for the preparation of the upper part of the osteotomy. As for the tap, this is resulting in no or very few bone fragments being created during implant placement.

UNIQUE PATENTED DESIGN INTENDED FOR BONE PRESERVATION*

The universality of Axiom X3® is the result of its patented thread design.

When manufacturing a classical implant, the threading tool makes a helical hollow with a defined pitch. But for Axiom X3®, a second threading cycle is performed with a slightly different pitch. By controlling the difference in pitch between the two threading cycles, it is possible to control the thickness of the thread along the entire implant body (Fig. 1). This results in a more progressive character of the insertion.



Patents #EP3763321B1 - EP376330B1




Figure 1 - Thread comparison between Axiom® PX and Axiom X3®, helical flutes removed

*Through a reduction in the number of drilling and tapping steps, compared to AXIOM® REG

In selected area of the external thread of Axiom X3®, represented in yellow in Figure 2, the top of the thread has been lowered by milling. Along the insertion, lowered and non-lowered thread portions are alternatively penetrating the bone.




Figure 2 - Axiom X3®

The stress to the bone and the friction is significantly reduced for those lowered zones.

“ THIS IS RESULTING, IN COMBINATION WITH THE PREVIOUS FEATURE, IN A BETTER CONTROL OF THE INSERTION TORQUE, ALLOWING INSERTION IN HARD BONE WITHOUT A BONE TAP ”

- HERV RICHARD, AXIOM X3 IMPLANT DESIGNER

Table 1 is showing in-vitro measured insertion torque of an Axiom X3® implant, Ø4.0 mm – length 12.0 mm, in foam blocks, representative of D1-type bone² (Sawbones, Pacific Research Laboratory Inc., USA), with a Ø3.6 mm drilled hole.

Table 1 - In vitro measured insertion torque

	PU BLOCK DENSITY (PCF)	IMPLANT BED DIAMETER (mm)	MAX. TORQUE (N.cm)	MEAN OF MAX. TORQUE (N.cm)
AXIOM X3® Ø4.0 - L 12.0 mm	40 <i>(representative of D1)</i>	3.6	34.6	37.9
			40.3	
			38.6	

*Data on file

As shown in Figure 3, even in hard bone situation, there is a remaining interference volume between the implant and the osteotomy. Therefore, bone fragments are formed during implant insertion and distributed around implant body.

The strains can therefore be distributed along the implant length and not concentrated at the neck (Fig. 4).




Figure 3 - External contour of Ø4.0 – L 12.0 mm Axiom X3® and shape of the Ø3.6 mm drill




Figure 4 - High resolution µCT evaluation of bone strains following insertion of a Ø4.0 – L 12.0 mm Axiom X3®[3]

During implant insertion, the top of the thread of the lowered zones are guided in the groove created by the previous non-lowered zone, which can facilitate implant placement.

“**AXIOM X3[®], IMPLANT DESIGN'S THINNER, ALTERNATING THREADS, HELPING TO IMPROVE SURGEON CONTROL.**
56% OF THE USERS HIGHLIGHTED IMPROVED GUIDANCE AS A MAIN FEATURE OF THE IMPLANT¹”

Anchoring the implant in the correct position in extraction socket is often tricky. The reduced size of Axiom X3[®] apex allows its insertion in narrow implant beds in soft bone. As an example, in very soft bone $\varnothing 4.0$ mm Axiom X3[®] implant can be inserted in a hole of only $\varnothing 2.0$ mm. **More bone is then preserved and the bone preparation duration is reduced (compared to AXIOM[®] REG)³¹.**

Thanks to the use of High Resolution Tomography (μ CT), we were able to measure bone densification, during in vitro experiments performed on domestic pig pelvis, and to characterize bone debris generated during implant placement of Axiom X3[®] (Fig. 5) and Axiom[®] REG (Fig. 6)³.

Bone debris are generated around Axiom X3[®] body.

Bone debris are generated around Axiom X3[®] body.




Figure 5 - Axiom X3[®] $\varnothing 4.0$ – L 12.0 mm in low density pelvic bone of domestic pig




Figure 6 - Axiom[®] REG $\varnothing 4.0$ – L 12.0 mm in low density pelvic bone of domestic pig

“**IT HAS BEEN SHOWN THAT THE PRESENCE OF AUTOGENOUS BONE FRAGMENTS CAN BE BENEFICIAL FOR THE REPAIR OF BONE⁷**”

Hervé RICHARD
R&D Engineer



“**AXIOM X3® WAS DEVELOPED FOR SURGEONS, WHO WANT TO HAVE AN IMPLANT THAT CAN BE PLACED IN DIFFERENT BONE DENSITIES AND IN MANY CLINICAL SITUATIONS, LOOKING FOR A SIMPLE AND READABLE PROTOCOL THAT DOES NOT REQUIRE THE FASTIDIOUS USE OF A TAP AND REDUCE OPERATING TIME**

”


Peri-implant bone densification was reported as an efficient mean to improve primary stability^{4,5,6}. The concept of Axiom X3® with its specific thread design and underpreparation is very straightforward to induce this local compaction in low-density bone, without additional tools nor specific procedures.

This densification was clearly visible after insertion in a low-density bone in Figure 5, where an Axiom X3® implant was surrounded by a shell of higher bone density.

The isolated bone fragments could be identified by image analysis all around the implant. In the same type of bone, whereas debris were

rare in a conventional protocol with Axiom® REG type implant there was a positive distribution of them along the Axiom X3® (Fig. 7 and 8)³.

It has been shown that the presence of autogenous bone fragments can be beneficial for the repair of bone⁷. In terms of primary stability, the μ CT observations are in good agreement with the reported implant stability in clinical practices, judged “well” or “excellent” in 82% of the cases in low density areas¹.



Figures 7 & 8 - Isolated bone debris around Axiom X3® (a) and Axiom® REG (b)

CONICAL CONNECTION, THE GOLDEN STANDARD FOR BONE LEVEL IMPLANTS

The connection of a Bone Level implant should ensure the tightness⁸ of the implant-abutment interface for crestal bone preservation. A Morse type connection (Fig. 9) allows an even distribution of mechanical constraints in the assembly and reduction of micro-movements at the implant-abutment interface⁹. **These features translate into the absence of bacterial infiltration at the interface level¹⁰ and allow the subcrestal placement of the implant, for easier aesthetic management.**

All Axiom[®] Bone Level implants have a conical Morse taper type connection of 6° half-angle.




Figure 9 - Conical Morse taper type connection of 6° half-angle

STRONG AND TIGHT CONICAL CONNECTION

AXIOM[®] BONE LEVEL SINGLE CONICAL CONNECTION - "ONE SIZE FITS ALL"

The single conical connection with triangular indexation offers flexibility and facilitates prosthetic management. The connection diameter is the same for all Axiom[®] Bone Level Ø3.4 to 6.4 implants, with X3, REG or PX profiles (Fig. 10).



Figure 10 - Topview of Axiom[®] connection

ABSENCE OF MICROGAPS

High-resolution X-Ray CT was used to test on an assembly built and loaded in accordance with the geometrical prescriptions of ISO 14801¹¹. The assembly was scanned first with no loading then under loading up to 150 N with 30° divergence (Fig. 11).

The tomography slices in the most loaded direction show **the absence of microgaps in the Axiom[®] BL connection under loading**, within the limit of the micron-size resolution.




Figure 11 - Tomography with no loading (a) and 150 N. loading (b)

THE CHOICE OF MECHANICAL PERFORMANCE

Inspired by its historical expertise in machining high strength elements for the spine or aerospace industry, Anthogyr selected the finest material to manufacture Axiom® implants. The implantable grade, extra-low-interstitials titanium alloy was selected. Well above the minimum strength requirements set in standards, the ultimate tensile strength of Grade V Titanium* bars used for Axiom® implants manufacturing often exceeds 1200MPa, due to its very fine microstructure (Fig. 12). Fatigue tests of Ø4.0 mm samples machined in the same conditions, confirmed the large difference in strength between Grade V Titanium* and Ti Grade IV on the long run. (Fig. 13)




Figure 12 - Microstructural analysis of Ti6Al4V-ELI in its delivery condition for dental implant machining[14]




Figure 13 - Compared fatigue resistance of identical specimen.

The raw material performance, maintained at a high level during machining and surface treatment steps, confers a very high fatigue resistance to the implants. **Mechanical resistance of narrow diameter implants made of titanium alloys like Grade V Titanium* is reported to be larger than that of Grade IV implants** in a laboratory study comparing various dental implant systems¹⁴. Another study with identical implant designs but Grade II and Grade V materials get to the same conclusion in terms of fatigue resistance¹⁶.

BCP SURFACE TREATMENT PROVEN CLEANLINESS

Anthogyr surface treatment relies on a blasting media BCP, Biphasic Calcium Phosphate, that is easily removed from the surface by acid treatment. Regular extensive analyses are performed to guarantee surface purity.

The surface analysis of 65 implant systems by scanning electron microscopy (Fig. 14 - 16) was performed by Dr Dirk Duddeck, Department for oral surgery and implantology, University of Köln, Germany on a large number of dental implants from many implant brands¹⁹.

“ THE RESULTS OF THIS STUDY ON THE AXIOM® PX IMPLANT DEMONSTRATED HIGH ACCURACY OF THE IMPLANT GEOMETRY AND CLEANLINESS AT THE MICRON SCALE, NO ORGANIC NOR INORGANIC CONTAMINANTS WERE FOUND ON THE SURFACE OF THE IMPLANT. ”




Figure 14




Figure 15




Figure 16

Figures 14 to 16 - Scanning Electron Microscope views of BCP surface treatment.

MODERATELY ROUGH SURFACE TOPOGRAPHY

Anthogyr Axiom® BCP implants were measured according to the guidelines defined by Wennerberg and Albrektsson²¹, and the classification proposed by the same authors²² was applied to qualify the surface. With a Sa value of 1.2µm, the BCP surface can be classified as moderately rough.

It was reported that moderately rough titanium surfaces like BCP showed stronger bone response than a smoother surface²³.

A profile based measurement according to metrology standard provides a corresponding Ra value between 1,5 and 2µm (Table 2). 3D surface analyses provide a more complete representation of the actual surface topography as illustrated in Figs 17 and 18.

Table 2 - Average 3D and 2D roughness of BCP implants.

3D MEASUREMENT (INTERFEROMETER) ACC TO ²	2D MEASUREMENT (PROFILOMETER) ACC. TO ISO4287
Sa, µm	Ra, µm
1.21(±0.12)	1.5-2




Figure 17 - Confocal 3D view of the BCP surface.




Figure 18 - Representative interferometric image of Anthogyr Axiom® implant surface.

BCP SURFACE TREATMENT IN VITRO BIOLOGICAL PERFORMANCE

Various in-vitro and preclinical studies established the performance of BCP surface treatments on titanium and titanium alloys²⁴⁻²⁷ and demonstrated **BCP high biological performance in terms of proliferation activity and ALP expression level of osteoblasts, which is favorable compared to gold-standard SLA-type surfaces²⁴.**




Figure 19

Figure 18 - Alkaline phosphatase (ALP) activity normalized to protein content of osteoblastic cells cultured on the titanium surfaces for 8, 15 and 21 days. Symbols indicate a statistical difference with $p < 0.05$ between the groups, adapted from²⁴.

IN VIVO BIOLOGICAL PERFORMANCE

BCP treated implants were recently compared with SLA-type surface implants, in an ovine model²⁸, with identical implant design and final cleaning/sterilization steps.




Figure 20 - Bone volume to Total volume ratio calculated from μ -CT measurements on SLA and BCP surface after 4 and 8 weeks of healing (internal data).




Figure 21 - Bone-to-Implant contact measured on SLA and BCP surfaces after 4 and 8 weeks of healing (internal data).

Bone remodeling kinetics were assessed by X-ray microtomography (μ -CT) and histology at 4 and 8 weeks, showing no statistically significant difference between BCP and SLA surfaces.

Histological analyses indicated a favorable healing scenario for BCP treated implants, from the early time point of 4 weeks, continued until reaching a BIC value of 79% at 8 weeks.

BCP SURFACE TREATMENT IN VIVO BIOLOGICAL PERFORMANCE

More recent implantation studies on Axiom® implants^{18,19} confirmed the osseointegration performance with Bone-to-Implant Contact (BIC) ratio reaching up to 68.4% at 4 weeks, and 92.0% at 13 weeks, obtained in representative animal models (dogs or minipigs) under Good Laboratory Practices. In a study by Chacun et al.¹⁷ Grade V Titanium* implants with an endosseous BCP surface were tested in a split-mouth design (each half-mandible receives either titanium or ceramic implants) in six Beagle dogs in a study aiming to evaluate the pre-clinical performance (local tissue effects and osseointegration properties) in comparison to ceramic implants, often referred as more favorable to biological integration.




Figure 22 - Microscopic view of an histologic cut on Axiom® implant.

Local tissue effects and BIC were evaluated at 4 and 13 weeks after implantation through histology.

Implant topographies were evaluated. Axiom® 2.8 implants presenting BCP, a moderately rough surface with a Sa value of 1.2μm, were inserted in the jaw bone and left for submerged healing.

After 4 weeks of healing, the BIC value for BCP reaches 68,4% (+/- 14,7%), and 92% (+/- 8,6%) at 13 weeks. **BIC values obtained for BCP are well above the threshold defined to expect sufficient bone anchorage over titanium implants*.**




Figure 23 - Mean (± SD) bone to implant contact at 4 and 13 weeks^[17].

*Considering Albrektsson report [29], who defined at 60% the threshold BIC value required to obtain sufficient bone anchorage over titanium implants.

CONCLUSION

Built on the experience of Axiom® PX in post-extraction sockets³⁰, Axiom X3® convinced its first users who used it in 32% of the cases for immediate treatments, with a satisfaction rate of 92% regarding its primary stability¹.

But immediacy does not limit to self-tapping implant designs and primary stability. For us it encompasses a wider ecosystem enabling an increased efficiency of surgeries and treatment plans. Axiom X3® integrates in the Axiom® system to render immediacy more universal with reduced protocols (no tap), but also wide diameters for wide post-extraction sockets, and comprehensive prosthetic range for immediate provisionalization or loading.


Wide diameters for post-extraction sockets



-25%

Straightforward protocols. Number of drilling steps reduced by 25% compared to Axiom® REG

Final prosthetic torque 25Ncm. Less risk of implant mobilization at prosthesis placement



“ **BUILT FROM THE ORIGINAL AXIOM® RANGE, RELYING ON THE THREE PILLARS OF MATERIAL, SURFACE, AND CONNECTION. AXIOM X3® SETS THE TONE FOR THE NEXT GENERATION OF IMPLANTS** ”

The Axiom® system has constantly been evolving for 15 years, providing our customers with an exhaustive solution to treat their patients with innovative and relevant products and workflows.

Thank to its design that has reached a very high level of maturity, our new Axiom X3® Bone Level implant is addressing a very large scope of clinical needs with minimal invasiveness and a straightforward user experience.

This leaves room for surgeons to refine and adapt protocols to the specific requirements of each clinical situation, either to maximize esthetic results or to apply immediate protocols in response to patient expectations, all within increasingly integrated digital workflows.

REFERENCES

- [1] Anthogyr. Axiom X3® international product launch survey - Report T22025, 2022.
- [2] Misch CE, Judy KW. Classification of partially edentulous arches for implant dentistry. *The International Journal of oral implantology implantologist* 1987;4(2):7–13.
- [3] Olivier C, Maire E, Richard H, et al. Bone behavior around a self-tapping dental implant during its insertion: European Synchrotron Radiation Facility 2025.
- [4] Lahens B, Neiva R, Tovar N, et al. Biomechanical and histologic basis of osseodensification drilling for endosteal implant placement in low density bone. An experimental study in sheep. *Journal of the Mechanical Behavior of Biomedical Materials* 2016;63:56–65.
- [5] Gaspar J, Proença L, Botelho J, et al. Implant Stability of Osseodensification Drilling Versus Conventional Surgical Technique: A Systematic Review. *Int J Oral Maxillofac Implants* 2021;36(6):1104–10.
- [6] Mercier F, Bartala M, Ella B. Evaluation of the Osseodensification Technique in Implant Primary Stability: Study on Cadavers. *Int J Oral Maxillofac Implants* 2022;37(3):593–600.
- [7] Shah FA, Palmquist A. Evidence that Osteocytes in Autogenous Bone Fragments can Repair Disrupted Canalicular Networks and Connect with Osteocytes in de novo Formed Bone on the Fragment Surface. *Calcified Tissue International* 2017;101(3):321–27.
- [8] Schmitt CM, Nogueira-Filho G, Tenenbaum HC, et al. Performance of conical abutment (Morse Taper) connection implants: A systematic review. *Journal of Biomedical Materials Research Part A* 2014;102(2):552–74.
- [9] Zipprich H, Weigl P, Lange B, et al. Micromovements at the implant-abutment interface: Measurement, causes, and consequences (in German). *Implantologie* 2007;1:31–46.
- [10] Assenza B, Tripodi D, Scarano A, et al. Bacterial leakage in implants with different implant-abutment connections: an in vitro study. *J Periodontol.*,2012, 83(4):491-7.
- [11] Adrien, J. et al (2016), Comparative in-situ loading analysis of narrow diameter implants by high resolution X-Ray tomography, Poster presented at the 25th EAO's Annual Scientific Congress, Paris, France.
- [12] ISO 5832-3. Medical devices — Implants for surgery — Metallic materials — Part 3: Wrought titanium 6-aluminium 4-vanadium alloy.
- [13] ASTM F136. Standard Specification for Wrought Titanium-6Aluminum-4Vanadium ELI (Extra Low Interstitial) Alloy for Surgical Implant Applications (UNS R56401).
- [14] Liens A, Fabregue D, Courtois N, et al. Multi-scale microstructural and mechanical characterization of titanium and titanium alloys used in dentistry. Poster presented at the 24th EAO's Annual Scientific Congress, Stockholm, Sweden. 2015.
- [15] Imam AY, Moshaverinia A, McGlumphy EA. Implant-abutment interface: A comparison of the ultimate force to failure among narrow-diameter implant systems. *Journal of Prosthetic Dentistry* 2014;112(2):136–42
- [16] Hirata R, Bonfante EA, Machado LS, et al. Mechanical Evaluation of Two Grades of Titanium Used in Implant Dentistry. *INT J ORAL MAXILLOFAC IMPLANTS* 2015;30(4):800–05.
- [17] Velasco-Ortega E, Jos A, Camean AM, et al. In vitro evaluation of cytotoxicity and genotoxicity of a commercial titanium alloy for dental implantology. *Mutation research* 2010;702(1):17–23.
- [18] Chacun D, Lafon A, Courtois N, et al. Histologic and histomorphometric evaluation of new zirconia-based ceramic dental implants: A preclinical study in dogs. *Dental materials official publication of the Academy of Dental Materials* 2021;37(9):1377–89.
- [19] Hoornaert A, Vidal L, Besnier R, et al. Biocompatibility and osseointegration of nanostructured titanium dental implants in minipigs. *Clinical oral implants research* 2020;31(6):526–35.
- [20] D.Duddeck, Department for oral surgery and implantology, University of Köln, Germany. Implant Study 2014 /2015 Quantitative and qualitative element analysis of implant surfaces by SEM/EDX.
- [21] Wennerberg A, Albrektsson T. Suggested guidelines for the topographic evaluation of implant surfaces. *The International journal of oral & maxillofacial implants* 2000;15(3):331–44.
- [22] Albrektsson T, Wennerberg A. Oral implant surfaces: Part 1--review focusing on topographic and chemical properties of different surfaces and in vivo responses to them. *International Journal of Prosthodontics* 2004;17(5):536–43.
- [23] Albrektsson T, Wennerberg A. Oral Implant Surfaces: Part2-Review Focusing on Clinical Knowledge of different surfaces. *International Journal of Prosthodontics* 2004;17:544–64.
- [24] Le Guehennec L, Lopez-Heredia M-A, Enkel B, et al. Osteoblastic cell behaviour on different titanium implant surfaces. *Acta biomaterialia* 2008;4(3):535–43.
- [25] Le Guehennec L, Goyenvalle E, Lopez-Heredia MA, et al. Histomorphometric analysis of the osseointegration of four different implant surfaces in the femoral epiphyses of rabbits. *Clinical oral implants research* 2008;19(11):1103–10.
- [26] Habibovic P, Li J, van der Valk CM, et al. Biological performance of uncoated and octacalcium phosphate-coated Ti6Al4V. *Biomaterials* 2005;26(1):23–36.
- [27] Shams N, Ghasemi M, Sadatmansouri S, et al. Morphology and differentiation of MG63 osteoblast cells on saliva contaminated implant surfaces. *Journal of Dentistry (Tehran, Iran)* 2015;12(6):424.
- [28] Anthogyr. Internal data.
- [29] Albrektsson T, Eriksson AR, Friberg B, et al. Histologic investigations on 33 retrieved Nobelpharma implants. *Clinical materials* 1993;12(1):1–9.
- [30] Chacun D, Gritsch K, Droguet G, et al. Retrospective cohort study of bone level implants up to 7years of follow-up: Poster presented at Europerio 2022. *Journal of Clinical Periodontology* 2022;49(523):143–288.
- [31] Anthogyr calculations on file. T21089 Removed bone volume REG vs PX vs X3.